


ICAR-National Bureau of Plant Genetic Resources
Pusa Campus, New Delhi – 110012.

ADDENDUM

With reference to an open tender notice for which the due date of opening is fixed on 30.08.2016, the specification for Camera 1 for Motorized Upright Microscope for Imaging of Coloured and Fluorescent specimens may be read as : **Minimum 1.3 mega pixel or better resolution with pixel Size 4.54 micro m x 4.54 micro or better.** All other specification, terms & condition will be remain same.


24/8/16
Asstt. Admin. Officer (P)
24/8/16